



LABMATE (ASIA) PVT. LTD.

BAID MEHTA COMPLEX, C- BLOCK, 1ST FLOOR,
 NO 16 (OLD NO.183), MOUNT ROAD, SAIDAPET, CHENNAI - 15. INDIA.
 TEL : 044-2220 0066 / 2220 0166
 E-mail : customersupport@labmateasia.com / sales@labmateasia.com
 Website : www.labmateasia.com

Customer Copy

Office Copy

INSTALLATION / SERVICE REPORT

Customer : <u>Indian Immunologicals limited</u>	Product : <u>BTx</u>
Address : <u>112-campus Gachibowli Hydrabad.</u>	Serial No. : <u>G-330139</u>
	Model : <u>BTx ECM630</u> S.W. Ver.
Contact Person : <u>Mr. Santhakumar</u>	Date(s) of Visit : <u>15/04/21</u>
Phone : _____ Fax : _____	Job Completed : Yes <input checked="" type="checkbox"/> No <input type="checkbox"/>
Department : <u>R&D</u>	
Reasons For Visit	
<input type="checkbox"/> Installation	<input type="checkbox"/> Warranty
<input type="checkbox"/> Free Service	<input type="checkbox"/> Inspection
<input type="checkbox"/> Service Contract	<input type="checkbox"/> Customer Training
<input type="checkbox"/> Paid Service	<input type="checkbox"/> Others

Description of Works Installation, Demonstration

- As per user confirmation visited to lab for the installation of ECM630.
- unpacked the box, and found that system reached safely.
- All the parts are received as per "PO" copy.
- checked all the input voltage & found satisfactory.
- Tested all the commands are checked and found o.k
- Tested with water, PBS, and plug was measured.
- Tester completed and found satisfactory.
- System is ready to use.

P. Santhakumar
 15/04/21

Sanjay Nayak
 15/04/21

Customer's Signature & Seal Date : 15/04/21

Engineer's Signature Date :

Customer Name : P. SANTHAKUMAR

Sanjay Nayak: 9940615563.