


INSTALLATION / SERVICE REPORTS

Customer Address: M/S: Indian Immunological Limited, Coity - 643007	 SMTR SCIENTIFIC INSTRUMENTS No.41, 4th cross, RamnjaneyaNagar, chikkalachandra, Uttarahalli, ! Bangalore-560061, KA, INDIA Mobile No.- +91-8825945722
---	---

EQUIPMENT DETAILS:

Instrument Name: SAS Air Sampler

Make/Model: SAS SUPER 100

SR.No: 21-D-16694/21-D-16699

Problems:


New SAS Air Sampler Installation


Action Taken:

Installation Activity completed successfully  
and Training session also completed. Instrument  
working normal condition

Customer Remark:

IQ, OQ, & PQ of Airsamplers Qualification Completed  
Software verification yet to be pending.

Engineer's Signature:   
Name: S. THARUNAL  
Designation: Sr. Service Engg  
Date: 31/01/2022

Customer's Signature:   
Name: P. O. J.  
Designation: Dy. Manager  
Date: 31.01.22