



Labtop Instruments Pvt. Ltd.
 Labtop House, Plot No. 59, Wally Phata, Sativalli Road,
 Vasai (E), Dist - Palghar - 401208, Maharashtra, India.
 T: +91-250-8457376-99 (24 lines)
 Email: info@labtopinstruments.com

SERVICE REPORT

SERVICE CALL / MAINTENANCE / INSTALLATION

Report No. **141**

Customer's Name & Address: Human Biologics Institute
Coty

AVL Date: 05/07/17
 AVL Time: 1:45 pm
 Dep. Time: 6:00 pm

Model: L 81300 + 400/6 Warranty Ends on _____
 A.M.C.

Machine SLR No.: 17026118

CALL DETAILS

1. Type of Calls: Break Down Maintenance Installation

Call Reported: _____ Call No. _____ Dt. _____

Nature of Calls Attended: The installation of the machine has been completed. The machine has opened at Set value of 43°C & 75°C.

Suggestions: _____

Details of Spares Replaced: Chargeable Free Replacement Under Warranty Under A.M.C.

1. _____ 4. _____
2. _____ 5. _____
3. _____ 6. _____

Customer Sign. & Seal: [Signature] 05.07.17

H.O.D.: [Signature] 05.07.17

Serviced by: Shy85 Sign. [Signature]



Labtop Instruments Pvt. Ltd.
 Labtop House, Plot No. 59, Waly Phata, Saitivali Road,
 Vasai (E), Dist - Palghar - 401208, Maharashtra, India.
 T: +91-250-6457376-99 (24 lines)
 Email: info@labtopinstruments.com

SERVICE REPORT

SERVICE CALL / MAINTENANCE / INSTALLATION

Report No. **142**

Customer's Name & Address: Humana Biological Institute
Coaly

Arvl. Date 6/17/12
 Arvl. Time 9:00 am
 Dep. Time 8:30 pm

Model: LB1-300T400/G Warranty A.M.C. Ends on _____

Machine SLR No.: 17026148 A.M.C.

CALL DETAILS

1. Type of Calls Break Down Maintenance Installation

Call Reported: _____ Call No. _____ Dt. _____

Nature of Calls Attended: The Validation for started for 24 hours of service of
43°C to 57.5°C for copper to down Chamber temp.

Suggestions: _____

Details of Spares Replaced: Chargeable Free Replacement Under Warranty Under A.M.C.

1. _____ 4. _____
2. _____ 5. _____
3. _____ 6. _____

Customer Sign. & Seal: [Signature] H.O.D. [Signature] Serviced by: Shylsh
 Sign. [Signature] 6/12/12

SERVICE REPORT

SERVICE CALL / MAINTENANCE / INSTALLATION

Report No. **143**

Customer's Name & Address

Human Biological Institute
Daly

Arvl. Date 7/7/17

Arvl. Time 9:00 am

Dep. Time 6:30 pm

Model

LB1-30044006

Warranty

A.M.C.

Ends on _____

Machine SLR No.:

170261148

CALL DETAILS

1. Type of Calls

Break Down

Maintenance

Installation

Call Reported

Call No.

Di.

Nature of Calls Attended

The Validation of the machine has been completed. The Software installation operation is also completed. Print outs have been taken & screenshots are taken.

Suggestions

Details of Spares Replaced

Chargeable

Free Replacement

UW
 UA

Under Warranty
Under A.M.C.

1. _____ 4. _____
2. _____ 5. _____
3. _____ 6. _____

Customer Sign. & Seal

[Signature]
07.07.17

H.O.D.

[Signature]
4.10.17

Serviced by: SHY/esh

Sign. *[Signature]*
7/7/17