

CUSTOMER SERVICE REPORT

SAMITEK INSTRUMENTS

1362, Sector - 16, Faridabad, Haryana - 121 002

Phone : 0129-4163032

Email : sales@samitekstruments.com

1124

No.

Date : 05 / 06 / 2020 Customer Principal Samitek

Reference No. PO NO 4511003199 — K. NAGARAJU

Service Nature Installation Warranty Paid Service Service Contractof Miscellaneous

Instrument : ...CAGE CHANGING STATION Customer : INDIAN IMMUNOLOGICALS
 Principal : ...TECHNIPART LIMITED
 Model / Type : ...9CS482JNNN RAKSHAPRAM
 Serial No. : ...18.00.2073 HYDERABAD - 500

Customer Equipment ID : N.A.
 Input Voltage : 232 V UPS Power RAW Power

Reported Problems : INSTALLATION

Actions taken : ...UNPACKED THE SHIPMENT... RECEIVED CAGE
 ...CHANGES STATION WITH GOOD CONDITION... SHIFTED IN ANAL
 ...ROOM POSITIONED AT INSTALLATION PLACE... PERFORMED
 ...THE CALIBRATION FOR LAF AIR VELOCITY AND EXHAUST
 ...AIR VELOCITY FOUND CALIBRATION WAS SUCCESSFUL... TESTED
 ...ALL WORKING FUNCTIONAL FOUND CAGE CHANGES STATION
 ...WORKING WITH GOOD CONDITION... PERFORMED DEMONSTRATION TO END USER

Engineer Comment / Follow - up action : ~~PROVIDE~~ UPS PROVIDE UPS POWER
 AND PLEASE GO THROUGH USER MANUAL BE FOR USE THE MACHINE

Spares Utilized : / / / /

Service Period On	From Hrs.	To Hrs.	Service Period On	From Hrs.	To Hrs.
05/06/2020	09:20	12:00	/	/	/

Customer Certification on Job Status Complete Incomplete

Customer Comments : Satisfied

Engineer Signature *[Signature]* 05/06/2020 Customer Signature with Date & Seal *[Signature]* 05/06/2020

Office Comments