



Indian Immunologicals Limited
KAMALAKAR REDDY
Rakshapuram, Gachibowli Post,
500032 HYDERABAD
INDIA

Service Report	
Number	1030021371
Date	22/05/2018
Customer reference	4511002520
Customer number	20013554
Inhouse contact	Technical Service BPS India
Phone	+91804357-3187
Fax	
Email	CustomerService.India@sartorius-stedim.com

Ship To:
Indian Immunologicals Limited
Rakshapuram, Gachibowli Post,
500032 HYDERABAD
INDIA

Bill To:
Indian Immunologicals Limited
Rakshapuram, Gachibowli Post,
500032 HYDERABAD
INDIA

First date of service: 11/05/2018
Last date of service: 11/05/2018
Services requested: Installation

Description of issue: Installation of Sartocheck 4+

Order Details

0010

Model	Serial Number	Inventory no.	Tag no.	
26288---FT	36201724			
Warranty Status	Cost center	Department	Building; Floor, Room	Site contact person
Out of Warranty				
Task	Installation (Status: Completed)	Technician: Krishna Mohan (40000080)		
51640	Task Status: Completed	Qty:		
Parts	-			
Labour	S8748INSTP - Installation Pack, FT/Sartocheck 4 Plus	1.00		
	S8799GTRAV - Travel costs general	2.00		
	87S99IWTA - Internal Working Time For Allocations	9.00		
	S8799GTRAV - Travel costs general	2.00		
	87S99IWTA - Internal Working Time For Allocations	7.00		
Technician Comment	1. Further to the previous visit, removed all accessories from packing and connected with the Sartocheck machine. 2. Performed the IQ/OQ as per the package received and handed over the documentation to the user. 3. Handed over all accessories to the user viz., INLET,OUTLET tubing, Network cable, Male stabuli connector and accessories. 4. System is working fine and is ready to use.			
Customer Comment	Thank you for your timely support.			

Date: 13 May 2018

Technician signature



Customer Signature

