

 techbio solutions® <i>your compliance partner</i>	CUSTOMER SERVICE REPORT ONSITE	Document No	TS/TE/FM-02B
		Issue No & Date	01 & 01-09-2015
		Rev No	2
		Issue Date	01-01-2022

CSR No.: TECH020601/CSR/23	DATE: 02-06-2023
-----------------------------------	-------------------------

COMPANY NAME: INDIAN IMMUNOLOGICALS LIMITED
--

COMPANY ADDRESS: RAKSHA PURAM, GACHIBOWLI POST, HYDERBAD - 500032

STATUS OF CALL: <input type="checkbox"/> Warranty/AMC <input checked="" type="checkbox"/> Installation <input type="checkbox"/> Calibration <input type="checkbox"/> Service	INSTRUCTION FROM: Mr N. SRINIVASA RAO
---	---

SERVICE REQUIRED / PROBLEM REPORTED

* Upgradation of LMS Express software onto new computer and IO qualification of new model Apex RSP onto existing location/system.



INSTRUMENT TYPE: <input checked="" type="checkbox"/> Particle Counter <input type="checkbox"/> Data Logger <input type="checkbox"/> Other	TYPE (for other only): MAKE: LIGHTHOUSE MODEL:
--	---

LOCATION OF INSTALLATION: B1 BLOCK	DATE: 02-06-2023	TIME: 15.30 Hrs
--	----------------------------	---------------------------

DETAILS OF SERVICE RENDERED

The upgradation of the software from LMS Express 7 to LMS Express 8 was completed successfully. The IO qualification of the model Apex RSP bearing S/N 2303154013, was successfully qualified onto the existing location/system.

CUSTOMER REMARKS:

TBS ENGINEER NAME: Prasath.T & Anuzuddin.RS	SIGNATURE WITH DATE:  02/06/2023  02/06/2023
EMP. CODE: TECH012 & TECH003	

CUSTOMER REPRESENTATIVE NAME: N. Srinivasa Rao.	SIGNATURE WITH DATE:  02.06.23
DESIGNATION: DY. Manager.	