

Hanna Equipments (India) Pvt. Ltd.

5/6, Aum Sai Building, Plot No 23C,
 Sector 7, Kharghar, Navi Mumbai - 410210.
 Phone: 022 27746555 / 68425211
 Email: tech@hannainst.in
 Website: www.hannainst.in

Service Tax Reg.: AACCH2670QST001
 VAT TIN : 27120729904
 GST ID Number : 27AACCH2670Q1Z4



INSTALLATION/ DEMONSTRATION/ SERVICE REPORT

Company: <u>M/ Indian Immunologicals</u>	Instrument: <u>PH/ORP/EC/TDS</u>	Report No.:
Contact Person: <u>P. Ramakrishna-Limited-TS</u>	Type: <u>Salinity</u>	Installation <input checked="" type="checkbox"/>
Department: <u>Engineering</u>	Model No.: <u>HI 5521-02</u>	Demonstration <input checked="" type="checkbox"/>
Address: <u>Karkapada</u>	Serial No.: <u>1231</u>	Warranty <input type="checkbox"/>
<u>Biotech park, phase-III</u>	S.No. <u>07460060101</u>	Out of Warranty <input type="checkbox"/>
<u>Siddipet</u>		AMC <input type="checkbox"/>
Phone: <u>9704546162</u>	Engineer: <u>D. Vijay</u>	No. of Days/Hrs.: <u>01 day</u>
Email: <u>p.ramakrishna@Indimmune.com</u>		
Date of visit: <u>17/11/2023</u>		

Nature of complaint: HI 5521-02 PH/ORP/EC/TDS/salinity (S.No. 07460060101)
Installation and Demonstration and IQ/OA/PA/DA qualification.
 Action taken: Initially checked Meter, Electrodes PH-HI1131B-S.No.094565EN,
EC/TDS - HI76312 - S.No. 095106BM. all are in good condition,
performed calibration of PH with 7.01PH, 4.01PH, 10.01PH
 Remarks: getting values are within the limit & satisfactory.
performed calibration of EC with 84uS/cm, 1413uS/cm getting values
are within the limit & satisfactory. IQ/OA/PA/DA documents qualification

Items replaced/ required

S. No.	Description	Quantity	Amount

Charges completed.

Spares charges: _____

Service charges: _____

Tax: _____

Total: _____

Customer remarks: Installation and demonstration has been completed
and. Date expiration demonstration need to be done.

Installation

Demonstration

Service

Work has been carried out and completed to our satisfaction

Customer sign with date: [Signature] 17-11-2023

Name of the Engineer: D. Vijay

Engineer sign with date: [Signature] 17/11/2023