



# SWAN ENVIRONMENTAL PVT. LTD.

An ISO 9001 : 2015 Certified Company

Plot No : 922&935, Swami Ayyappa Co-op Society, Madhapur, Hyderabad, Telangana - 500 081. India.

Office : 040-40216184/85. Fax : 040-40216183 Email: service@swanbiotec.com

## SHIMADZU TOC Analyzer INSTALLATION COMPLETION REPORT

SL.No.: 001

Customer P.O.No.	4511002268	Date of Installation	14/06/2018
Invoice No.	SL1050885	Distributor P.O.No.	26/SB/Q017
Sales Note	SI-1029195-H	Distributor Name	Swan Environmental Pvt. Ltd.
Company Address	Indian Immunologicals Ltd., Human Biologicals Ltd., Kozhikannal, Padumud Post, udaga mandalam, 643007, TN.		
Customer Dept/Email		Mobile Number	
Instrument Model	<input checked="" type="checkbox"/> TOC L <sub>CPH</sub> <input type="checkbox"/> TOC L <sub>CSH</sub>	Serial Number	TOC-L: H54435501823
	<input type="checkbox"/> TOC L <sub>CPH</sub> <input type="checkbox"/> TOC L <sub>CSH</sub>		TOC 4200: NA
	<input type="checkbox"/> TOC 4200 <input type="checkbox"/> Other : _____		OCT-L: NA
	<b>Accessories</b>		ASI-L H57415501845
	<input checked="" type="checkbox"/> ASI-L <input type="checkbox"/> OCT-L		SSM-5000: NA
<input type="checkbox"/> TNM-L <input type="checkbox"/> SSM	OTHER: NA		
<b>Check Points</b>		<b>Yes</b>	<b>No</b>
<b>Short Shipment Items : If Yes, Please Mention Items below:</b>			
1.		<input type="checkbox"/>	<input checked="" type="checkbox"/>
2.			
Any External damage Observed		<input type="checkbox"/>	<input checked="" type="checkbox"/>
Installation Environment meets with the specification		<input checked="" type="checkbox"/>	<input type="checkbox"/>
Instrument performance by standard IQ, OQ & PQ		<input checked="" type="checkbox"/>	<input type="checkbox"/>
Operational and Maintenance trainings are completed by engineer.		<input checked="" type="checkbox"/>	<input type="checkbox"/>
Lab Solution Manager DB Installed		<input type="checkbox"/>	<input checked="" type="checkbox"/>
lab Solution Server CS Installed		<input type="checkbox"/>	<input checked="" type="checkbox"/>
PM or Service agreement are explained by engineer		<input checked="" type="checkbox"/>	<input type="checkbox"/>
Software and Firmware Version	Software Version : 1.06    Firmware Version : 1.09    Lab Solution CS / DB : NA.		
Engineer Comments			
Customer Comments			
Engineer Name & Designation	J. Sivasankar Manager Service	Customer Name & Designation	V. K. RAMESH EXECUTIVE
Engineer Signature	<i>J. Sivasankar</i>	Signature	<i>V. K. Ramesh</i>

**SUN TEKNOLOZY / SUN LABORATORY SERVICES / SUN SCIENTIFIC CORPN**

14/1 Lakshmi Dwarka Villa, Thalayari Street, West Mambalam, Chennai-600 033, Ph.:044-24740143, Fax : 044-42024721  
 Email: admin@sunteknozy.com, sunteknozy@gmail.com

**SERVICE REPORT  
 CUSTOMER DETAILS**

12/06/18, 13/06/18 &  
 Date: 14/06/18

Customer Name & Address	Contact Person	Mr. D. Venkatesan	
Human Biologicals Institute	Designation	Manager	
Indian Immunologicals Ltd./	Dept.:	Quality Control	
Kozhipannai, Pudumund Post	Tel. No.:	O.:	Extn.: NA
Udhagamandalam - 643007	Mobile.:	90107 22266	
Tamilnadu	E-Mail.:	NA	

**INSTRUMENTS DETAILS**

Instrument Name	TOC Analyser with Autosampler
Instrument Model	TOC LCPH + ASI
Sr. No	H54435201823, H57415201842

**SERVICE DETAILS**

Installation   
  warranty   
  AMC   
  Services   
  Calibration  
 Chargeable/Free. if chargeable Amount Rs. \_\_\_\_\_

**Problem Reported:**  
 New Toc & Autosampler installation.

**Action Taken:** Supplied all the items as per the purchase order. Connected UHP Nitrogen cylinder (Inox) with TOC analyser. Performed IQ, OQ, PQ for TOC analyser. Given training on software operation. Informed about the Do's and Don't in the TOC analyser. Hence now completed the TOC installation. TOC & ASI are working in good condition & performed one sample analysis.  
**WARRANTY:** 12 months from the date of installation or 18 months from the date of supply (whichever is earlier) - Warranty does not cover for any glassware, battery and consumables.

**SPARE PARTS REPLACED / RECOMMENDED**

Sr. No.	Item Code	Item	Qty.	Replaced	Recommended
1		Purification Kit	1		✓
2		SS Regulators for cylinder	1		✓
		NA			

**Customer's Remark:** Instrument installed successfully and IQ, OQ, PQ performed.

J. Sivarajkumar 14/06/18 J. Sivarajkumar Engineer's Signature	Machine Working Satisfactory 14-06-18 Customer's Seal & Signature
------------------------------------------------------------------------	-------------------------------------------------------------------------