



TRIDENT EQUIPMENTS PVT. LTD.

A - 105,, Kailas Industrial Complex, Hiranandani Godrej Link Road, Vikhroli (W),
Mumbai - 79. Tel.: 2518 1704 / 05 / 06 Telefax : 2518 5463 E-mail : info@tridentequipments.com

Service Report

Name of Customer: <u>Indian Immunologicals Ltd</u>	Product: <u>Checkpoint Tec</u>
<u>Human Biological Institute</u>	Model(s): <u>check point</u>
Address: <u>Korhipanni, Budhmand (PO), ucha</u>	Serial No.(s): <u>1053</u>
<u>Gamandolan, cery 643007</u>	Installation <input checked="" type="checkbox"/> Warranty <input checked="" type="checkbox"/> Chargeable <input type="checkbox"/> Courtesy <input type="checkbox"/>
Contact Person/Dept: <u>Mr. R.H. Singh, P. Nagarajulu</u>	Repaired at - Customer Site <input checked="" type="checkbox"/> Service Centre <input type="checkbox"/>
PO No.:	AMC : Preventive Maintenance Visit No. _____
Telephone:	Breakdown Visit No. _____
Email: <u>rh.singh@indimmunol.com</u>	Date: <u>26 & 27/03/15</u>
Nature of Problem:	

Intubation.

Action Taken:

- ① Installed the Instrument and left it for flashing over night
- ② Tried to upgrade the firmware but found hanging Problem
- ③ checked 4 times found same Problem.
- ④ Finally found there is a problem in firm ware updation and brought back the Instrument for checking Purpose.

Engineer's Comment:

Instrument brought back along with me for checking & with latest firmware.

Customer's Comment:

Rectification the problem as early as possible and bring it back install very earliest.

Service Engineer's Name & Signature:

[Signature]

Customer's Name, Signature & Stamp:

[Signature]
27/03/15

Encl.: