

INSTALLATION REPORT

CUSTOMER NAME	INDIAN IMMUNOLOGICALS LTD (1208272)	INSTALLATION DATE	31-Aug-2021
ADDRESS	RAKSHAPURAM	CASE NO	167810
	GACHIBOWLI POST	MODEL NO	PRECISION BATH
CITY	HYDERABAD	STATE	TG
PINCODE	500 032	SERIAL NO	300443031
CONTACT PERSON / DESIG. / DEPT.	B.Lakshmi Narayana Yadav / NA / NA	OTHER DETAILS OF THE UNIT	
TELEPHONE / FAX	919959050521	INSTRUMENT STATUS	Warranty
EMAIL	b.lakshminarayana@indimmune.com	CALL REASON	Installation
		DIVISION	LSI



CALL DETAILS	
---------------------	--

DETAILS OF SERVICE RENDERED	Unpacked the consignment and verified as per PO. Carried out installation as per instruction manual/pir/checklist. Checked the performance of the machine, working OK. Run the machine with different set point temperatures and found working OK. Performed the IQ, OQ and PQ of the machine successfully. Installation and demonstration was not satisfactory.
------------------------------------	---

REMARKS (IF ANY)	Machine is working fine.
-------------------------	--------------------------

JOB COMPLETED	Yes
----------------------	-----

VISIT DETAILS			
DATE	FROM	To	HRS. WORKED
31-Aug-2021	11:33 AM	05:17 PM	5.73

	
CUSTOMER FULL NAME / SIGNATURE STAMP	Engineer's NAME / Signature

CUSTOMER PO NUMBER	378(269)/2020/IIL/TFS-M/GBS	INVOICE NUMBER	9240315355
SALES ORDER NUMBER	6606184		