

**PERFORMANCE QUALIFICATION**



Client : Indian Immunologicals, Ltd,  
 Location: Hyderabad.  
 Manufacturer: Microflow Devices India Pvt Ltd  
 Item/Job Name: Vertical Laminar Airflow Unit  
 Model no: MFD-V(PCR)-610  
 Sr.No: 3651/02/2022

**8.0 Summary and Conclusion:**

The Performance Qualification is carried out and the Parameters like Air velocity, Pressure and Air borne Particle Count tests were performed and the results were within limits and found ok. The reports were attached for reference.

|               |                           |                            |                   |
|---------------|---------------------------|----------------------------|-------------------|
| Signature     | Ravi                      | <br>B.V.V. Eswar           | <br>K. Thirath    |
| Date          | 17/06/22                  | 20/06/22                   | 20/06/22          |
| Name          | Ravi                      | B.V.V. Eswar               | K. Thirath        |
| Department    | Customer Support Engineer | Engineering                | Quality Assurance |
| Reviewed By : | <br>Vignat<br>20/06/22    | <br>K. Thirath<br>20.06.22 |                   |