

INSTALLATION REPORT

End User's Name	Mr. A.S.R. Pavan			
Designation	Executive	Deptt.	Production	
Company	Indian Immunology Limited			
Address	Rakshapavan		City	Hyderabad
	Aschibevli post		Pin-code	500022
	State	Telangana	Mob.	
Contact No.				
e-mail Id	a.pavan@indimmun.com			

Date of Visit	04/11/2019	Reference	
Type of Call	Pre-Installation	Installation	<input checked="" type="checkbox"/>

Instrument Details	
Type Of Product	PH meter
Catalog No.	15944-90
Model	PH 700
Sr. No	2584826

Initial Check list		
Damage to Instrument	Yes	No <input checked="" type="checkbox"/>
Damage to parts	Yes	No <input checked="" type="checkbox"/>
Part Missing (Check against packing list)	Yes	No <input checked="" type="checkbox"/>



Description: Installation and demonstration of PH meter.


Any additional comments: Instrument along with standard accessories received in good condition. Instrument installed. IQA checklist given. performance checked. Demo given on operation and calibration.

Status Of Job	Completed <input checked="" type="checkbox"/>	Pending
---------------	---	---------

Engineer's Remark: Instrument is working fine.

Customer's Remark:

Engineer's Name: Mr. Rajashkar Potla	End-User's Name: A.S.R. Pavan
Signature: 	Signature: 
04/11/2019	



INSTALLATION REPORT

End User's Name	Mr A.S.R. Pawar			
Designation	Executive	Deptt.	Production	
Company	Indian Immunology Limited			
Address	Roshapuram		City	Hyderabad
	Gachibowli post			
	State	Telangana	Pin-code	500032
Contact No.			Mob.	8790256144
e-mail Id	a.pawar@indimmunol.com			

Date of Visit	04/11/2019	Reference	
Type of Call	Pre-Installation	Installation	<input checked="" type="checkbox"/>

Instrument Details	
Type Of Product	pH meter
Catalog No.	1594-4-90
Model	PH 700
Sr. No	2884841

Initial Check list		
Damage to Instrument	Yes	No <input checked="" type="checkbox"/>
Damage to parts	Yes	No <input checked="" type="checkbox"/>
Part Missing (Check against packing list)	Yes	No <input checked="" type="checkbox"/>

Description: Installation and demonstration of pH meter.

Any additional comments: Instrument along with standard accessories received in good condition. Instrument installed. IQA checklist given. Performance checked. Demo given on operation and calibration.

Status Of Job	Completed <input checked="" type="checkbox"/>	Pending
---------------	---	---------

Engineer's Remark:	Instrument is working fine
Customer's Remark:	

Engineer's Name: Mr. Rajasekar. potu	End-User's Name: A.S.R. Pawar
Signature:	Signature:

04/11/2019
Cole-Parmer India Pvt. Ltd.
403, A-Wing, Delphi, Hiranandani Business Park, Powai, Mumbai 400 076
Tel: 91-22-61394444 Fax: 91-22-61394422 # E-Mail: service@coleparmer.in