



Lyophilization Systems India Pvt Ltd

SERVICE REPORT

Doc. No: LSIPL/EC-F-001 Rev-00

Customer Information

Complain Information

Customer Name : IIL
 Complained by : Mr. Phani Kumar
 Phone No : 9440373734

Model No: Lyodryer Project No. : 740
 Complain Date :
 Complain Number :

Terms: Warranty AMC AMC - Break Down Free Visit Paid Visit

Complain Details

Operation Parameters : Lab Units

Nucleation chamber and camera installation

Stabilizer Power Supply : 205V
 PLC Voltage : 5.1V
 Discharge Pressure 1 : 220 Pa^{SP}
 Discharge Pressure 2 : 110 Pa^{SP}
 Fluid Pressure : —
 Fluid Level : 30%
 Insulation Status : OK
 Shelf Operation : OK
 Shelf temp : —
 Cond Temp : -80.6°
 Vacuum and Oil : OK

Cause Identified

- do -

Call Completion Information

* Condenser temp trail taken +20°C TO -40°C reached in 5mins
 * Condenser ultimate temp is -80.6°C
 * Nucleation cycle trail taken and all parameters found satisfactory.

Spares/Parts Used :

Call Status :

Completed:

Not Completed:

Attended Date : 13/16/17/09/2019

Attended Time: 10:30 AM

Attend by : ch. chandrasekar

Customer's Sign & Stamp

Service Engg Sign:

[Signature]
17/09/2019

Customer's Remarks: