

SERVICE REPORT

Our Ref. :	Ubc / Reddy / 1129 / 2011
Customer	Endura Engineering & IT
Address	Gachibowli Hyderabad.
Scientist	Mr. Sunitha
Phone	— Fax — Telex —
S.No.:	65365H
Model :	RS 6100.K2.M
Instrument Details :	Precision
Invoice Details :	1129
Engineer's Report No.	1129
Principals	Reddy
Activity	
Installation	<input checked="" type="checkbox"/>
Warranty	<input checked="" type="checkbox"/>
Contract	<input type="checkbox"/>
Out of Warranty	<input type="checkbox"/>
Billable	<input type="checkbox"/>

Work Details : Emulsion?

• Installed the balance in laboratory.
 • Done the internal/external calibration
 • Done the 20, 0.0g and 100 qualification with reference
 • Submitted the all certificates in format return
 • Done the DPO to end user.
 • New balance is working in good condition

PARTS REPLACED

Part Number	Description	Quantity	Source	Amount
/	/	/	/	/

The service requested by us has been handled by O. Nageswara Babu
 Who visited us on 07/04/2011
 He has carried out the above work to our satisfaction.
 Customer Report :

Place : Gachibowli Date : 07/04/2011
 Engineer Signature : [Signature]
 Customer Signature : [Signature] (with Seal)
07.04.11