

TOWA OPTICS (INDIA) PVT. LTD.

Flat No. 402, Devi Residency Lane No. 2, Sai Enclave Colony,
 Habsiduja, Hyderabad - 500 007 Phone : 040-64593168
 E-mail : towahyd@towaoptics.com Web. : www.towaoptics.com
 CIN NO. : U74899DL1963PTC003902
 GSTIN No. : 36AAACT0010G1Z2

Eg. Id: B9/cms/134;

Location: Cell culture S. No.161.....

Date ..10-11-2021

INSTALLATION / SERVICE REPORT

User Details :

| | | | | |
|-----------|---|---------------------------------------|------------|--|
| User : | Name | Indian Immunological Ltd Rajeshapuram | | |
| | Department | Grachibowli Post Hyderabad. | | |
| | Contact No. | 9490065900 | Email ID : | |
| Address : | Indian Immunological Ltd Grachibowli Hyderabad. | | | |

Installation Service Warranty AMC / CMC FOC Service Visit

| | | | |
|---|---|---|---|
| 1 | 2 | 3 | 4 |
|---|---|---|---|

Concerned Sales person for FOC : D. Rameesh Kumar 9849992028

Commercial Details :

| | | | |
|-----------------------|---------------------------|---------|-------------|
| Order No. | 4513002603 TWINX/576/2020 | Dated : | AUG 10-2021 |
| Product received from | Nikon JAPAN | | |
| Invoice No. : | KKM 3181 | Dated : | AUG 10-2021 |

Product Configuration :

| No. | Item | Description | | | |
|-----|--------------|--------------------------------------|---------------------|--------------------|------------|
| 1 | Microscope | Model : Ci-L | S. No. 719091 | Ph. : 10x20x40x100 | DIC : — |
| | | Objective : 10x 20x 40x 100x 100xph. | NAMC : | | |
| 2 | Fluorescence | Model : — | S. No. — | Lamp :W | S. No. : — |
| | | Filters : — | — | | |
| 3 | Camera | Model : — | S. No. — | — | |
| | | Controller : — | S. No. — | — | |
| 4 | Software | Version : — | S. No. — | — | |
| 5 | Computer | Model : — | CPU / Laptop S. No. | TFT S. No. | |
| | | Printer S. No. : — | UPS Model — | S. No. | |
| 6 | | — | — | — | |
| 7 | | — | — | — | |

Engineer Comments :

| | |
|-----------|---|
| Work Done | checked the Ci-L microscope Accessories and Installed as per the order - checked with samples working condition |
|-----------|---|

Customer's Comments :

| |
|---|
| Installation and demo completed |
| Installed - working condition satisfactory. |

This is certify that Instrument Ci-L microscope Installed as per the order. Installation completed with samples demonstrated. The equipment is working to our entire satisfaction. Condition

10-11-21

D. Rameesh Kumar

Customer Signature with rubber stamp

Engineer Name & Signature

Warranty : Standard warranty one year or as per purchase order. No warranty of consumable or Third party items.

Further up - gradation possibility :