

**INSTALLATION COMPLETION REPORT**

DATE: 18/10/2022

CUSTOMER NAME: M/S Indian Immunologicals Limited

PERSON CONTACTED: Mr. Phani Kumar      DESIGNATION: Sr. Executive - Engg

MACHINE: Autoclave      JOB NO: 20200353

INSTALLATION STARTED ON: 29-09-2022      INSTALLATION COMPLETED ON: 18-10-2022

OPERATING MANUAL:       HANDED OVER       TO BE SUPPLIED

SHORT SUPPLIED IF ANY:

Nil

DAMAGED IF ANY: Plant steam PRV damaged condition. need replace new with one.

MALFUNCTIONING IF ANY:

Nil

ARE YOU SATISFIED WITH INSTALLATION?      HAPPY      QUITE OK      OK      NOT SATISFY

ARE YOU SATISFIED WITH OUR ENGINEER?      HAPPY      QUITE OK      OK      NOT SATISFY



DO YOU HAVE ANY SUGGESTION PLEASE WRITE:

Kindly Replace the Damaged PRV.

**PLANNING FOR COMMISSIONING**

UTILITY DETAIL:      TENTATIVE AVAILABILITY DATES / WEEKS

1. BOILER STEAM	18/10/2022
2. D.M.WATER	18/10/2022
3. COOLING WATER	18/10/2022
4. ELECTRICAL	18/10/2022
5. AIR SUPPLY	18/10/2022
6. VACUUM	18/10/2022

FOR	DATE:	DATE: 18/10/2022
NAME D PHANI KUMAR	 SIGN	 FOR Pharmalab INDIA PVT LTD.
DESIGNATION Sr. Executive		