



Complaint No.: —	Service Report No.: 9516
Complaint Date : —	Service Date : 04/11/2020
Customer Name : Indian Immunologists/H.	Contact Person : B. Murali Krishna
Address : Sy. No. 281-284 & 321, Kalkapatta Village, Telangana.	Designation : Chief Manager.
	Tel.:
	E-mail Id :
Equipment Name : walk in BOD Incubator	Sr. No.: 207/19-20 Id No.: KHOJ/WI/005
Reason For Call :	
Installation <input checked="" type="checkbox"/>	Breakdown <input type="checkbox"/> Calibration <input type="checkbox"/> AMC <input type="checkbox"/> Other <input type="checkbox"/>
Complaint Details : Regarding Installation of walk in BOD Incubator	
Action Taken : IQ, DQ, PLC validation, OQ, software Installation & PQ with 1 hour Calibration & 24 hour with load and without load validation are been performed.	
Observation : Installation of chamber as well as of software is done. Chamber is working properly according to set values	
Material Replaced : <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No	Job Completed <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No.
1) <del>NA</del>	Pending work : <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No.
2) <del>NA</del>	
3) <del>NA</del>	
4) <del>NA</del>	
5) <del>NA</del>	
Customer Remark :	
For MACK PHARMATECH PVT. LTD.	
Signature : <i>Akshay Bhale</i>	For Customer
Name : Akshay Bhale	Signature : <i>T. Mithun Kumar</i>
Designation : Service Engineer	Name : T. Mithun Kumar
	Designation : Executive.
<b>Office Use Only</b>	
Under warranty : <input type="checkbox"/> Yes <input type="checkbox"/> No	Under AMC : <input type="checkbox"/> Yes <input type="checkbox"/> No
W.O./No.:	P.O./No.:
Chargeable : <input type="checkbox"/> Yes <input type="checkbox"/> No. Invoice No.:	Date
	Amount :