




SERVICE REPORT

| | | |
|--|---|--|
|  <p>MIG - 315, APHB Colony, Balaji Nagar, Kukatpally, Hyderabad - 500 072. Cell : (0) 9505555984 7097092726 Email : krishna@tecpharma.in</p> | Date : 27-02-2024 | S/R: 23441 |
| | Engineer's Name : D. KRISHNA | |
| | Mobile No. : 7097092726 | |
| Customer's Name : | Type of Services | |
| Address Indian Immunologicals Ltd | 1. Installation <input checked="" type="checkbox"/> | 6. Paid Service <input type="checkbox"/> |
| Gachibowli | 2. Breakdown <input type="checkbox"/> | 7. Calibration <input type="checkbox"/> |
| Hyderabad, | 3. AMC <input type="checkbox"/> | 8. IQ/OQ/PQ <input checked="" type="checkbox"/> |
| | 4. Warranty <input checked="" type="checkbox"/> | 9. Application <input type="checkbox"/> |
| | 5. Inspection <input type="checkbox"/> | 10. Validation <input type="checkbox"/> |
| P.O.No. 4511004653 | Job Completed : Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> | |
| E-mail : | | |
| Instrument osmo meter | Model No. ISMO TOUCH 2 | Sr. No. / INST ID 684990141/ A5/osm/583 |
| Problem As Reported by the Customer : Osmo meter Installation. | | |
| Service Done : Installed the osmometer and done the IQ, OQ and PQ andification. During PQ verified the standards and results are within acceptance limit. operational training given to the user's. | | |
| Note : Warranty three years from today. | | |
| Spare Parts Used / Required | To Be filled by Customer | |
| NA | <input checked="" type="checkbox"/> Instrument Installed & Found Working Satisfactorily <input type="checkbox"/> Prevention Maintenance Carried Out Satisfactorily <input type="checkbox"/> Fault Rectified & Instrument Working Satisfactorily <input type="checkbox"/> Please await our written approval for parts replacement <input type="checkbox"/> Calibration done Satisfactorily <input type="checkbox"/> Comments (If Any) | |
| Engineer's Signature :  | Name : V. Talun | Customer Signature & Date  27-02-24 |
| | Phone Number : 8121304306 | |
| | * Only for parts Changed & all other charges will be extra as applicable | |

Note : Parts replaced are chargeable except during warranty period. Consumables used for the instrument are not covered by warranty & hence are chargeable. Parts replaced due to mishandling during operation or otherwise will also be charged in every case. In case of any difficulty or suggestion, please contact Service Manager.