

INSTALLATION REPORT

End user's Name	Dr. D. Vengatesan.		
Designation	Deputy Manager	Deptt.	Quality control
Company	M/S Human Biological Institute		
Address	Ooty		
			City Ooty
	State	Tamilnadu	Pin Code 643007
Contact No.			Mob. 8012755566
e-mail Id	vengatesan@indimmune.com		

Date of Visit	21.02.17	Reference	
Type of Call	Installation	<input checked="" type="checkbox"/>	Pre-Installation <input checked="" type="checkbox"/>

Instrument Details	Categories	
Instrument of	Cole-Parmer	
Type Of Product	pH meter	Printer
Catalog No.	35420-20	15944-89
Model No	pH 2700	SRP350
Sr. No	2568015	MAX4AKA16080113


Initial Check list		
Damage to Instrument	Yes	No <input checked="" type="checkbox"/>
Damage to parts	Yes	No <input checked="" type="checkbox"/>
Part Missing (Check against packing list)	Yes	No <input checked="" type="checkbox"/>

Description:
Installation and Demonstration.
Supplied along with pH electrode, stand, pH buffer bottles of 4,7 and 10, electrode refilling solution, Manuel in CD format and hard copy, electrode manual as hard copy, ELECTRODE TEST REPORT, RS232 CABLE and calibration certificates.

Any additional comments:
Installed, demonstrated and calibrated the pH meters. Calibrated with 4, 7 and 10 pH buffer solutions. Explained electrode storage and maintenance procedures. Performed IQ/OQ and handed over the document.

Status of Job	Completed <input checked="" type="checkbox"/>	-----
---------------	---	-------

Engineer's Remark: pH meter is installed, calibrated, demonstrated and qualified to customer's satisfaction. Submitted the IQ/OQ documents.
Customer's Remark:

Engineer's Name: Rajesh. V	Customer's Name: Dr. D. Vengatesan.
Signature: 	signature: 