

<b>End user's Name</b>	Mr. V.S.R. Anjaneyulu		
<b>Designation</b>	Manager	<b>Dept.</b>	Production
<b>Company</b>	Indian Immunologicals Limited		
<b>Address</b>	Rakshapuram, Gachibowli Post		
	Hyderabad	<b>City</b>	Hyderabad
	<b>State</b>	Telangana	<b>Pin-code</b> 500033
<b>Contact No.</b>	-----	<b>Mob.</b>	9666250099
<b>e-mail Id</b>	anjaneyulu@indimmune.com		

<b>Date of Visit</b>	10/02/2020	<b>Reference</b>	4511002966
<b>Type of Call</b>	Pre-Installation	<b>Installation</b>	✓

<b>Type Of Product</b>	Benchtop Meter
<b>Catalog No.</b>	35420-20
<b>Model</b>	Oakton PH2700
<b>Sr. No</b>	2909158

Initial Check list		
<b>Damage to Instrument</b>	Yes	No ✓
<b>Damage to parts</b>	Yes	No ✓
<b>Part Missing (Check against packing list)</b>	Yes	No ✓

**Description:**
**Installation and demonstration:**

Benchtop meter PH2700 is Supplied along with power supply adapter, Installation-Operation Manual (C.D), Quick guide, pH electrode (2907020), Temperature probe, pH buffer solutions, Electrode refilling solution, Thermal Printer with power supply adapter and communication cables.

**Any additional comments:**

Installed the instrument along with printer.  
 Calibrated the instrument with 3 point buffer solutions and slope (4.01-7.00: 98.0% & 7.00-10.01 : 96.8%) found pass.  
 Installation completed successfully.

<b>Status of Job</b>	Completed	Yes ✓	Completed	No
----------------------	-----------	-------	-----------	----

**Engineer's Remark:**

Benchtop Meter PH2700 is installed and demonstrated to customer's satisfaction.

**Customer's Remark:**

<b>Engineer's Name:</b> Shreekantha K Nayak	<b>End-user's Name:</b> Mr. V.S.R. Anjaneyulu
<b>Signature:</b> 	<b>Signature:</b>  10-02-20

**Cole-Parmer India Pvt. Ltd.**

403, A -Wing, Delphi, Hiranandani Business Park, Powai, Mumbai 400 076  
 Tel: 91-22-61394444 Fax: 91-22-61394422 # E-Mail: service@coleparmer.in