

# Hanna Equipments (India) Pvt. Ltd.

3/4/5/6, Aum Sai Building, Plot No. 23C, Sector 7, Kharghar, Navi Mumbai - 410210.  
 Tel: +91-22-27746554 / 55 / 56, Fax: +91-22-27746557  
 Email: tech@hanna-india.com, Website: www.hannainst.com/usa  
 Service Tax Regn: AACCH2670QST001 VAT TIN: 27120729904



## INSTALLATION/ DEMONSTRATION/ SERVICE REPORT

Company: <u>INDIAN IMMUNOLOGICAL PVT LTD</u>	Instrument: <u>HI5221-02</u>	Report No.: <u>2073</u>
Contact Person: <u>Mr. Neelkantam</u>	Type: <u>ph meter</u>	Installation <input checked="" type="checkbox"/>
Department: <u>Research and development</u>	Model No.: <u>HI5221-02</u>	Demonstration <input checked="" type="checkbox"/>
Address: <u>Gaeehibowli</u>	Serial No.: <u>F0106499</u>	Warranty <input type="checkbox"/>
		Out of Warranty <input type="checkbox"/>
Tel: _____		AMC <input type="checkbox"/>
E-mail: _____		

Date of visit: <u>20/2/2018</u>	Engineer: <u>J. FAYAZ</u>	No. of Days/Hrs.: <u>1</u>
---------------------------------	---------------------------	----------------------------

Nature of complaint: Installation of HI5221-02 calibration of 7, 4, 10 done and with two probes HI1313 and HI1083 successfully.

Action taken: Demonstrated and explained with 2 probes along with calibrations and samples also done

Remarks: Installation done successfully.

Parts replaced/ required				Charges	
S. No.	Description	Quantity	Amount	Spares charges:	_____
				Service charges:	_____
				Tax:	_____
				Total:	_____

Customer remarks: Installation and train was done satisfactorily. PH Meter calibration Certificate to be provided with due date for both the equipments installed at ILL RD. This was explained to service engineer during installation. Dtg. 20.02.2018.

The Demonstration <input checked="" type="checkbox"/> Installation <input type="checkbox"/> Service <input type="checkbox"/> Has been carried out and completed to our satisfaction  <u>Neelkantam</u> Customer sign with date: <u>20/02/18</u>	Name of the Engineer: <u>J. Fayaz</u>  <u>[Signature]</u> Engineer Sign with date: _____
--	--