

# Hanna Equipments (India) Pvt. Ltd.

3/4/5/6, Aum Sai Building, Plot No. 23C, Sector 7, Kharghar, Navi Mumbai - 410210.

Tel: +91-22-27746554 / 55 / 56, Fax: +91-22-27746557

Email: tech@hanna-india.com, Website: www.hannainst.com/usa

Service Tax Regn: AACCH2670QST001 VAT TIN: 27120729904



## INSTALLATION/ DEMONSTRATION/ SERVICE REPORT

Company: <u>INDIAN IMMUNOLOGICAL LTD.</u>	Instrument: <u>HI 5221-02</u>	Report No.: <u>2072</u>
Contact Person: _____	Type: <u>PH/MU Meter</u>	Installation <input checked="" type="checkbox"/>
Department: <u>RESEARCH AND DEVELOPMENT</u>	Model No.: <u>HI 5221-02</u>	Demonstration <input checked="" type="checkbox"/>
Address: <u>TIL</u>	Serial No.: <u>F0107096</u>	Warranty <input checked="" type="checkbox"/>
<u>GACCHIBOWLI, HYDERABAD</u>		Out of Warranty <input type="checkbox"/>
Tel: _____		AMC <input type="checkbox"/>
E-mail: _____		

Date of visit: <u>19/2/2018</u>	Engineer: <u>J. FAYAZ</u>	No. of Days/Hrs.: <u>1</u>
---------------------------------	---------------------------	----------------------------


Nature of complaint: Installation of ph/MU meter HI 5221-02


Action taken: calibration with 7, 4, 10 points done. along with MU readings also checked sample with 7.36 measured.

Remarks: demonstrated and calibrated with sample readings accepted.

Parts replaced/ required				Charges	
S. No.	Description	Quantity	Amount	Spares charges:	_____
				Service charges:	_____
				Tax:	_____
				Total:	_____

Customer remarks: Installation and train was done satisfactorily.

The  
 Demonstration   
 Installation   
 Service   
 Has been carried out and completed to our satisfaction  
 19/2/18  
 Customer sign with date:

Name of the Engineer:  
J. FAYAZ  
  
 Engineer sign with date: