

INSTALLATION / SERVICE REPORT



ULTRA INSTRUMENTS Plot No. 92, Amarjyothi Colony, Bowenpally, Secunderabad - 500 011. Tel: (9140) -27951055, 27951066 Email: us@ultragroup.in		Ref. No.: 283 Customer Service Order No.:	
Customer : <u>IPC</u> <u>Graph + Software</u>		Installation <input checked="" type="checkbox"/> Breakdown <input type="checkbox"/> P.M.S. <input type="checkbox"/> General Visit <input type="checkbox"/>	
Contact : <u>V. Srinivas</u> Phone/Mobile : <u>9705350025</u>		Inv No. <u>USI/1512/10-21</u> Dt. <u>29-SEP-2020</u> PO.No. <u>4511003261</u>	
E-mail ID : <u>vadla.srinivas@indimmm.com</u>		Installation Date <u>30-3-2021</u>	
Product : <u>PH Meter</u> Sl.No. : <u>20310016</u> Make : <u>OHMUS</u> Car / Model <u>ST 5000</u>	Remarks : <u>Installation</u> <u>Instrumentation handed</u> <u>over to customer in</u> <u>good condition. 229 V.A.C</u>		
Diagnosis / Action Taken <u>Installation of ST-5000 PH meter</u> <u>is completed. Training also given to user</u> <u>PH-meter calibration done with PH 7 & 11.4</u>			
Completed <input checked="" type="checkbox"/>		Repeat Visit <input type="checkbox"/> <u>Buffer needed found ok</u>	
ENGINEER'S REMARKS		CUSTOMER'S REMARKS	
<u>Instrument working</u>			
<u>Setter done</u>			
Engineer's Signature <u>[Signature]</u> Engineer's Name <u>V. Srinivas</u>		Customer's Signature <u>[Signature]</u> With Seal <u>[Seal]</u> 3003.21	