

INSTALLATION REPORT

End user's Name	Mr. Vadla Srinivas		
Designation	Manager	Deptt.	HH production
Company	M/S Indian Immunologicals Limited		
Address	Rakshapuram, Gachibowli post		
		City	Hyderabad
	State	Telangana	Pin Code 500032
Contact No.		Mob.	970535225
e-mail Id	vadlasrinivas@indimmune.com		

Date of Visit	19.10.16	Reference	
Type of Call	Installation	<input checked="" type="checkbox"/>	Pre-Installation <input checked="" type="checkbox"/>

Instrument Details	Categories			
Instrument of	Cole-Parmer			
Type Of Product	pH meter	pH meter	pH meter	pH meter
Catalog No.	58825-24	58825-24	58825-24	58825-24
Model No	A211	A211	A211	A211
Sr. No	X33041	X33046	X33047	X32563

Initial Check list		
Damage to Instrument	Yes	No <input checked="" type="checkbox"/>
Damage to parts	Yes	No <input checked="" type="checkbox"/>
Part Missing (Check against packing list)	Yes	No <input checked="" type="checkbox"/>

Description:
Installation and Demonstration.
Supplied along with pH electrodes, stands, pH buffer bottles of 4,7 and 10, electrode refilling solution, Manuel in CD format and hard copy, electrode manual as hard copy and calibration certificates-1 sets each and data cables of 2 each

Any additional comments:
Installed, demonstrated and calibrated the pH meters. Calibrated with 4, 7 and 10 pH buffer solutions. Explained electrode storage and maintenance procedures.

Status of Job	Completed <input checked="" type="checkbox"/>	-----
---------------	---	-------

Engineer's Remark:
4 Nos pH meters are installed, calibrated, demonstrated and qualified to customer's satisfaction. Submitted the IQ/OQ documents.

Customer's Remark:
Installation completed

Engineer's Name: Rajesh. V	Customer's Name: Mr. Vadla Srinivas
Signature: <i>Rajesh</i> 19/10/16	signature: <i>Vadla</i> 19.10.16