



TRIDENT EQUIPMENTS PVT. LTD.

A - 105,, Kailas Industrial Complex, Hiranandani Godrej Link Road, Vikhroli (W),
Mumbai - 79. Tel.: 2518 1704 / 05 / 06 Telefax : 2518 5463 E-mail : info@tridentequipments.com

Service Report

No. **63633**

Name of Customer : <u>Indian Immunological Limited</u>	Product : <u>GE TSC Sensor</u>
Address : <u>Korakalade</u>	Model(s) : <u>CHECK POINT</u>
Contact Person/Dept : <u>Mr. Ramakrishna Eng.</u>	Serial No.(s) : <u>23033991</u>
PO No. : <u>4511004457</u>	Installation <input type="checkbox"/> Warranty <input checked="" type="checkbox"/> Chargeable <input type="checkbox"/> Courtesy <input type="checkbox"/>
Telephone : <u>-</u>	Repaired at - Customer Site <input checked="" type="checkbox"/> Service Centre <input type="checkbox"/>
Email : <u>-</u>	AMC : Preventive Maintenance Visit No. <u>NA</u>
	Breakdown Visit No. <u>NA</u>
	Date : <u>17-11-2023</u>

Nature of Problem : Performance Qualification (PQ)

Action Taken :

→ After IQ, OQ Activities, Control the Investment in water plant found ok

→ As per PQ Performed the LOD of Log Test found all values are within the limit as per Specification.

Engineer's Comment : PQ completed Successfully.

Customer's Comment : ok

Service Engineer's Name & Signature : Mahesh A. 17/11/2023

Customer's Name, Signature & Stamp : P. Rama Krishna 17.11.23