

# SERVICE REPORT



Date : 24/11/2020

S/R : 1612

Engineer's Name : D. Ashok

Mobile No. : 8897885376

G-315, Balaji Nagar, Kukatpally, Hyderabad - 500 072,  
Ph (O) : 9505555984, E-mail : krishna@tecpharma.in

Customer's Name :

Address Indian Immunological Limited  
Rakshapuram, Gachibowli post  
Hyderabad - 500032  
Telangana, India

**Type of Services**

- |                 |                                     |                 |                                     |
|-----------------|-------------------------------------|-----------------|-------------------------------------|
| 1. Installation | <input checked="" type="checkbox"/> | 6. Paid Service | <input type="checkbox"/>            |
| 2. Breakdown    | <input type="checkbox"/>            | 7. Calibration  | <input type="checkbox"/>            |
| 3. AMC          | <input type="checkbox"/>            | 8. IQ/OQ/PQ     | <input checked="" type="checkbox"/> |
| 4. Warranty     | <input type="checkbox"/>            | 9. Application  | <input type="checkbox"/>            |
| 5. Inspection   | <input type="checkbox"/>            | 10. Goodwill    | <input type="checkbox"/>            |

Instrument No.

Job Completed : Yes  No

Instrument Model No.

Model No.

Sr. No. / INST ID

Instrument Name: Haier Ultra Low Temperature Freezer DW-86L729BP  
Freezer

Sr. No. / INST ID: BE06Y8EBA000GLP0007/  
B5/ULF/220

Problem As Reported by the Customer : Haier ULT Freezer Installation

Service Done : Haier Ultra Low Temperature Freezer Installation, Completed.

Performed IQ, OQ and PQ Qualification during 72 hours temperature mapping without load, 72 hours temperature mapping with load at -70°C ±10°C, power failure study and door open study done.

Result: now the Ultra Low Temperature freezer working satisfactorily

**Spare Parts Used / Required**

NA

**To Be filled by Customer**

- Instrument Installed & Found Working Satisfactorily
- Prevention Maintenance Carried Out Satisfactorily
- Fault Rectified & Instrument Working Satisfactorily
- Please await our written approval for parts replacement
- Calibration done Satisfactorily
- Comments (If Any)

Name : P. Subraman

*[Signature]*

Date : 24-11-20

Customer Signature

Engineer's Signature : D. Ashok 24/11/2020

\* Only for parts Changed & all other charges will be extra as applicable

Notes : Parts replaced are chargeable except during warranty period. Consumables used for the instrument are not covered by warranty & hence are chargeable. Parts replaced due to mishandling during operation or otherwise will also be charged in any case. In case of any difficulty or suggestion, please contact Service Manager.