

CL# _____ Date 24-04-2019

- AHMEDABAD
 BANGALORE
 CHANDIGARH
 CHENNAI
 DELHI
 HYDERABAD
 KOLKATA
 LUCKNOW
 MUMBAI
 PUNE

Institution: Indian Immunologicals Limited, Karakapatta.
 Attn: Mr. Arun / Mr. Srinivas L
 Department: HHQC
 Tel: +91 9553331100
 Email: j.arun@indimmune.com

Order Ref: 4513002248 dated: 04-02-2019.
 Model: 50T5 8
 Serial No: 190318C
 Base code: 1550200
 Version: 1.05
 Assaycode: _____
 Version: _____
 Principals: Biotek
 NuAire
 UVP
 Promical
 Biospherix

Service Complaint: Installation of Biotek 50T5 washer
 Observation: _____
 Service Details: _____
 Utilities Check List: Electric Supply : 230VAC
 Ground / Earthing
 Voltage Stabiliser / UPS
 Air Conditioning
 Dust Free Environment
 Computer
 Printer
 Interface : Serial / USB
 COM Port: _____
 Gas Supply
 Distilled Water

Service Details: Commissioning completed and calibration completed. 10-00-PA documents completed and submitted.
 Material Used: _____
 Part No: _____
 Description: _____
 Qty: _____
 Invoice V/N: _____
 Test Result: The instrument installed and qualification also completed. Working satisfactorily.
 Note / Remarks: _____

For MEDISPEC (I) LTD
 Engineers Name: A. Jateegh
24-04-2019
 Customers Approval / Acceptance: _____
 Signature & Seal: H. Srinivas
24-04-19